

Circitrend V5 Recorder

Model Selection Guide

Instructions

- Select the desired key number. The arrow to the right marks the selection available.
- Make the desired selections from Table I using the column below the proper arrow.
A dot (•) denotes unrestricted availability. Restrictions follow Table VII.

Key Number I II III IV V VI VII

_____ - _____ - _____ - _____ - _____ - _____ - _____ - _____

KEY NUMBER	Selection	Availability
Circitrend V5 Electronic Data Recorder	TVCP	↓

TABLE I - ANALOG INPUTS

Analog Inputs	Two	2 _	•
	Four	4 _	•
	Six	6 _	•
	Eight	8 _	•
Input Type	Standard	_ 0	•
	Universal Inputs	_ U	•

TABLE II - DISCRETE INPUTS/OUTPUTS & ANALOG OUTPUTS

Discrete Inputs/ Outputs	None	0 _	•
	Four Relay Outputs	4 _	•
	Eight Relay Outputs/Two Digital Inputs	8 _	•
	Eight Relay Outputs/Eight Digital Inputs	A _	•
No Selection	None	_ 0	•

TABLE III - FIRMWARE OPTIONS

Firmware	None	0 _ _	•
	Math + Totalizers + Event Markers + Groups	A _ _	•
No Selection	None	_ 0 _	•
No Selection	None	_ _ 0	•

TABLE IV - COMMUNICATIONS

Communication Protocol	None	0 _ _	•
	Trendbus RS485 (Real time)	T _ _	•
	Modbus	M _ _	•
Instrument Power	90 - 240 VAC	_ 0 _	•
	24 VDC	_ 2 _	•
Transmitter Power	None	_ _ 0	•
	Transmitter Power	_ _ P	c

TABLE V - DATA STORAGE/MEMORY CARD**Selection Availability**

Data Storage	1.44MB Floppy	0 _ _	•
No Selection	None	_ 0 _	•
No Selection	None	_ _ 0	•

TABLE VI - OPTIONS

Case/Mounting	Standard Panel Mounting	0 _ _ _ _	•
	NEMA4X/IP65 Enclosure with Latch & Wall Mount Kit	M _ _ _ _	•
	NEMA4X/IP65 with Key Lock & Wall Mount Kit	R _ _ _ _	•
Documentation	English Manual with TrendViewer	_ U _ _ _	•
	Product Information on CD ROM with TrendViewer	_ 0 _ _ _	•
Tagging	None	_ _ 0 _ _	•
	Linen Tag (Note 1)	_ _ L _ _	•
	Stainless Steel Tag (Note 1)	_ _ S _ _	•
Approvals	None (CE Standard)	_ _ _ 0 _	•
Certifications	None	_ _ _ _ 0 _	•
	Certificate of Conformance (F3391)	_ _ _ _ B _	•
	Custom Calibration Test Report (Note 2)	_ _ _ _ C _	•
	Cert. of Conformance & Calib. Test Report (Note 2)	_ _ _ _ E _	•
Software (Note 3)	None	_ _ _ _ _ 0	•
	Trend Manager Pro	_ _ _ _ _ P	•
	Trend Server Pro (Single User License)	_ _ _ _ _ S	•

TABLE VII

Factory Use Only		00	•
------------------	--	----	---

RESTRICTIONS

Restriction Letter	Available Only With		Not Available With	
	Table	Selection	Table	Selection
c			I	_ U

- Notes:**
1. Customer must supply tagging information, Up to 3 lines of 22 characters each are allowed.
 2. Customer must supply Input Actuation and Range information for each input in the free form section of the order, otherwise the calibration certificate will be based on the factory default ranges.
 3. Software Packages can be ordered separately (see Accessories price page).